

QI-ROG



Visit the QI-ROG page
for news, updates and downloads



CONTENTS

Product overview.....	3
Technical specifications.....	4
Electrical characteristics.....	4
Sensitivity in conductor placement.....	4
Mechanical characteristics.....	4
Environmental characteristics.....	4
Safety.....	4



SAFETY WARNINGS AND CAUTIONS

The following warnings and cautions must be observed to ensure personal safety and prevent damage.



Death or **serious injury** may result from failure to heed this warning.



It is necessary to comply with national regulations when installing and picking materials for power lines.



Material damage or **serious personal injury** may result from failure to heed this warning.



Repairs and modifications must be carried out only by the manufacturer. It is forbidden to open the case and make any changes to the device. Tampering with the device will invalidate the warranty.



The manufacturer **declines all responsibility** for electrical safety in the event of improper use of the equipment.



It is essential to read the entire contents of this manual before carrying out any work.



The product described in this document may only be used for the specified application. The maximum performance data and environmental conditions specified in the product data sheet must be observed. Proper transport and storage, as well as professional assembly, installation, handling and maintenance are required for the correct and safe operation of the device.



Installation and commissioning must be carried out by qualified personnel only.

Before commissioning, make sure that:

- the maximum values for all connections are not exceeded; refer to the product data sheet;
- the connection cables are not damaged or live during wiring;
- the direction of current flow and phase rotation are correct.

During installation, ensure that a switch or circuit-breaker is near the product and easily accessible.

The unit must be uninstalled if safe operation can no longer be guaranteed (e.g. visible damage). Disconnect all connections in this case. The unit should be returned to the manufacturer or to an authorised service centre for repair.

Use under ambient conditions other than those specified, application of signals or voltages other than those specified, may cause significant deviations from the specified measurement tolerances, which may be irreversible.



WARNING: High-intensity magnetic fields may alter the values measured by the transformer. Avoid installation near: permanent magnets, electromagnets, or iron masses. If irregularities are detected, reposition or move the unit to a more suitable location.



Although the contents of this document have been checked for accuracy, it may contain errors or inconsistencies and we cannot guarantee its completeness or accuracy.



This document is subject to periodic revision and updating. QEED reserves the right to make changes to the product and/or its technical documentation at any time in the interests of continuous quality improvement. Always consult the latest version of the documentation available on the website:

www.qeed.it

If you find any errors or missing information in this document, please notify us by e-mail to:

technical@qeed.it



Failure to observe the warnings may result in damage to the equipment or failure to operate as intended.



Disposal of waste electrical and electronic equipment (applicable in the European Union and other countries with separate collection). The symbol on the product or its packaging indicates that the product should not be treated as household waste. Instead, it will be handed over to an authorised collection point for the recycling of electrical and electronic waste. Ensuring that the product is disposed of properly will prevent potential negative effects on the environment and human health, which could otherwise be caused by inappropriate waste management of the product. Recycling materials helps to conserve natural resources. For further information, please contact your local authority, waste disposal service or the retailer from whom you purchased the product.



Please note that the information on the nameplate must be observed.





PRODUCT OVERVIEW

Rogowski QI-ROG probes are suitable for measuring alternating and pulsed currents. The flexible and lightweight head allows installation in hard-to-reach areas and on large conductors. Available in various lengths and colours.

The probes can be used with our **QE-POWER-x** and **QC-ENERGY-3T** power meters or with our **QE-CURRENT-485** signal converter.




Figure 1: QI-ROG dimensions



TECHNICAL SPECIFICATIONS

Electrical characteristics

Max. measurable current	100 kA @ 50 Hz
Accuracy	± 1%
Linearity	± 0,2%
Output signal	100 mV / 1000 A @ 50 Hz
Working frequency	20 Hz... 5 kHz

Sensitivity in conductor placement

Conductor	± 2% maximum from closing point
External field influence	± 0,5% max
Temperature sensitivity	± 0,07% by °C

Mechanical characteristics

Probe cross-section diameter (C)	12 mm
Length of open probe (B)	300, 400, 500, 600, 700 mm (more on request)
Outer diameter of locking connector (D)	17 mm (max)
Maximum diameter of the conductor or bar to be measured (A)	QI-ROG-300: 84 mm QI-ROG-400: 115 mm QI-ROG-500: 147 mm
Materials	Probe and cable: Thermoplastic rubber, UL94 V-0
Joints	PA6 UL94 V-0
Cable	1000V UL STYLE 20940; outer diameter 5mm; conductor 2x 26 AWG
Cable length	2, 4, 10 m (based on the model)
Shielding	100% coil, 100% output cable

Environmental characteristics

Working temperature	-20°C ÷ +70 °C
Relative humidity	85% max without condensation
Pollution degree	2
Maximum operating height	2000 m

Safety

Max. working voltage	1000 V @ 50/60 Hz
Overvoltage category (IEC 61010-1)	CAT III
Hi-pot test (sensor and output cable)	7400 Vac @ 50/60 Hz for a minute